


Product Data Sheet

Precision alcoholometer

Reference 0920FC060/20-qp

- t With fiducial mark
- t With Individual number



Reading below meniscus



Specifications

- t Glass alcoholometer - Ethanol %Vol
- t Scale : 60-80 %Vol
- t Division : 0.2 %Vol
- t Accuracy: ± 0.2 %Vol
- t Calibration temperature 20°C
- t Surface tension : 27.9 mN/m
- t Reading below meniscus
- t ± 335 mm
- t With a Certificate of Conformity
- t Traceable Calibration Certificate on request
- t In individual box

Alla France

Z.I. du Bompas - 13, rue du Bompas - Chemillé - 49120 Chemillé-en-Anjou - FRANCE
Tél. + 33 (0)2 41 30 55 08 - Fax +33 (0)2 41 30 34 67

www.allafrance.com

