



Flasks, volumetric, class A, USP/ISO/DIN tolerances **PYREX®**

- Complies with ISO 1042 and DIN 12664
- Calibrated to the tolerance required by the United States Pharmacopoeia (USP) which is tighter than those of ISO/DIN standards
- Each flask is laser etched with an individual serial number, batch code and date of manufacture
- Batch certificate available to download online
- Conformity 'H', certified
- Manufactured from chemically resistant Pyrex® glass
- Inscriptions in white enamel
- Polyethylene stopper

Product Code	Capacity (ml)	Tol. (+/- ml)	Approx height (mm)	Stopper size	Pack qty.
3151/01	1	0.020	65	7/16	5
3151/02	2	0.020	65	7/16	5
3151/04	5	0.020	70	10/19	5
3151/06	10	0.020	90	10/19	5
3151/08	20	0.03	110	10/19	5
3151/10	25	0.03	110	10/19	5
3151/12	50	0.05	140	12/21	5
3151/14	100	0.08	170	14/23	5
3151/16	200	0.10	210	14/23	5
3151/18	250	0.12	220	14/23	2
3151/20	500	0.20	260	19/26	2
3151/22	1000	0.30	300	24/29	2
3151/24	2000	0.50	370	29/32	1
3151/26	5000	1.00	475	34/35	1



20ml and 5000ml sizes are supplementary to DIN range. Please note that the downloadable batch certificates can be obtained by visiting www.scilabware.com and entering the unique batch code, which can be found on the base of each flask.
UKAS certified volumetric flasks are available on request.