


Volumetric Glassware: Measuring Pipettes, Graduated



GVP-PRA776AS

Various Types of Graduated Pipettes

Type 1

- Zero at top to any graduation line down to shoulder

Type 2

- Zero at jet

Type 3

- Zero at top of jet

Type 4

- Zero at jet with the last drop to blowout (i.e. serological + blowout pipettes)

Measuring Pipettes, Graduated, Type 1, Class AS

- Complies with ISO 835.
- Calibrated to deliver from zero at the top to any graduation line down to the shoulder.
- Graduation in permanent amber stain which resists aggressive washing solutions.

Brand	Reference	Cap/mL	Std/Pack
LMS	GVP-63336161	0.1 x 0.001	10
LMS	GVP-63336162	0.2 x 0.002	10
LMS	GVP-63336163	0.5 x 0.01	10
MBL	GVP-PRA770AS	1 x 0.01	10
MBL	GVP-PRA772AS	2 x 0.02	10
MBL	GVP-PRA774AS	5 x 0.05	10
MBL	GVP-PRA776AS	10 x 0.10	10
LMS	GVP-63336169	20 x 0.10	10
MBL	GVP-PRA778AS	25 x 0.20	10
LMS	GVP-63336182	50 x 0.20	10

Measuring Pipettes, Graduated, Type 1, Work Cert, Class AS

- Complies with ISO 835.
- Calibrated to deliver from zero at the top to any graduation line down to the shoulder.
- Graduation in permanent amber stain which resists aggressive washing solutions.
- Each pipette is supplied with calibration certificate showing actual volume dispensed and providing traceability to national standards.

Brand	Reference	Cap/mL	Std/Pack
MBL	GVP-PRA780ASWC	1 x 0.01	10
MBL	GVP-PRA782ASWC	2 x 0.02	10
MBL	GVP-PRA784ASWC	5 x 0.05	10
MBL	GVP-PRA786ASWC	10 x 0.10	10
MBL	GVP-PRA788ASWC	25 x 0.20	10

Measuring Pipettes, Graduated, Type 1, Class B

- Complies with ISO 835.
- Calibrated to deliver from zero at the top to any graduation line down to the shoulder.
- Graduation in permanent amber stain which resists aggressive washing solutions.

Brand	Reference	Cap/mL	Std/Pack
Clearance Items (while stocks last)			
Class B			
HBG	GVP-NR/816B	0.2 x 0.002	10
HBG	GVP-1000142	2 x 0.02	10